



Verification Assay Result Sheet

Verification Assay For Microbiology Laboratory Technicians
(Ref. [MICLAB 105](#))

Verification Assay Date: _____

Name of Technician: _____

Test Reagents

Reagents	Lot No.	Reconstitution Date	Expiry date	
Pyrogen				EU/mL sensitivity
Endotoxin				EU/mL potency
Pyrosperser		NA		2% working concentration
Test kit		NA		

L.A.L. Endotoxin & Endotoxin Working Standards diluent.

Any sterile batch (WFI) (Tested to be L.A.L. negative) Batch No.: _____ Expiry: _____

Test Session Standards - Results

Key: (+) firm gel, (-) no gel or viscous gel.

Replicate Assay Number	Endotoxin Concentration & Gelation Results (EU/mL)							Endpoint	
	1	0.5	0.25	0.125	0.06	0.03	0.015	EU/mL	Log ₁₀
1.									
2.									
3.									
4.									

Negative Controls

Key: (+) firm gel, (-) no gel or viscous gel.

Replicate Assay No.	Control Results
1	
2	



Verification Assay Result Sheet

Verification Assay For Microbiology Laboratory Technicians
(Ref. MICLAB 105)

NEGATIVE CONTROL/S

Calculations:

SDy =

$$\sqrt{\frac{n \left(\sum_{i=1}^n Y_i^2 \right) - \left(\sum_{i=1}^n Y_i \right)^2}{n (n - 1)}}$$

SDy =

$$\sqrt{\frac{(\quad) - (\quad)}{(\quad - 1)}}$$

=

Where Yi = log10 Xi (i = 1,2...n)
Xi = Gelation end-point in EU/mL
And n = number of replicate end-point assays

Standard: 99 % upper limit on SD Log₁₀ x (Lysate sensitivity) where n = 4 is 0.365.

SDy = _____

Has the standard been met? YES/NO



Verification Assay Result Sheet

Verification Assay For Microbiology Laboratory Technicians
(Ref. [MICLAB 105](#))

Calculations:

$$GMx = 10^{\check{y}}$$

where $\check{y} =$

$$\frac{\left(\sum_{i=1}^n y_i \right)}{n}$$

()

where $\check{y} =$ _____

=

$$GMx = 10^{\check{y}}$$

=

=

Standard: GM sensitivity value should fall within the range 0.03 – 1.25 EU/mL.

GMx = _____ EU/mL

Has the standard been met? YES/NO

Comments about Session: _____

Signature of Technician:	Approved by:
Date:	Date: