

VAL 200: Selection and Use of Biological Indicators during Validation Studies

Appendix 2: Biological Indicator Test Results Sheet

Project Number	Equipment ID

Procedure

Fill out table and transfer sheet to Quality Control with all samples.

Cycle Number:		Run #:		Initials/Date:		
QC Serial Number:						
Commenced Incubation at		°C	Time:		Initials/Date:	

1. To activate the media, place the indicator in an upright position in a plastic crusher. Gently squeeze the crusher to break the glass ampoule. Place the activated indicator in the incubator rack and incubate immediately.
2. Examine the indicator for any colour change at 24h. Record the examining time and results in the table.

Acceptance Criteria - *Geobacillus stearothermophilus*

- Yellow colour indicates bacterial growth.
- No colour change indicates adequate sterilisation.

Position	Op. Init.	Date/Time	Biological Indicator Purple = - Yellow = + 24h	Initial/Date
Control				
Operator Initial/Date				
Test Result:		<input type="checkbox"/> Pass		<input type="checkbox"/> Fail
Comment:				
Tested By/Date				
Verified By/Date				